

West Gippsland Waterwatch Volunteer Water Quality Monitoring Data Sheet

Site Details	
Site Code:	HAZ 017
Site Name:	HAZEL CK at Ascott SE
Site Grid Reference Easting:	406187
Northing:	5776729

Test Details	
Date:	16 SEPT '10
Time:	1.00 <input type="checkbox"/> AM <input checked="" type="checkbox"/> PM
Person(s) Conducting the Survey / Test:	DAVID
Name of Monitoring Group:	WARRAGUL LANDCARE

RECEIVED

21 SEP 2010

BY: _____

Your Meter - Reading Before Calibration (when in CAL Mode)				Your Meter - Reading Post Calibration (after pressing the ENT button)			
EC Meter	1400	pH Meter in 4.01	4.2	EC Meter	1360	pH Meter in 4.01	4.6
pH Meter in 7.01	7.5	pH Meter in 4.01	4.2	pH Meter in 7.01	7.1	pH Meter in 4.01	4.6

Water Condition (Circle)					
Steady Flow	Rising	Peak	Falling	Stagnant (pools)	Dry (no water)

Test Results		
Air Temperature	13.4	°C
Water Temperature	12.0	°C
pH	5.5	
Electrical Conductivity	210	<input type="checkbox"/> µS <input type="checkbox"/> mS
Reactive Phosphorus	0.015	mg/L P
Turbidity	< 10	N.T.U.
Dissolved Oxygen:		<input type="checkbox"/> mg/L <input type="checkbox"/> %

Any Comments for Your Facilitator?

Weather Conditions at Time of Sampling	
<input type="checkbox"/> Sunny	<input type="checkbox"/> Cloudy
<input checked="" type="checkbox"/> Overcast	<input type="checkbox"/> Raining
<input checked="" type="checkbox"/> Windy	

Water Appearance	
<input checked="" type="checkbox"/> Clear	<input type="checkbox"/> Muddy
<input type="checkbox"/> Milky	<input type="checkbox"/> Foamy / Frothy
<input type="checkbox"/> Stained Green	<input type="checkbox"/> Stained Brown
<input type="checkbox"/> Smelly	<input type="checkbox"/> Scummy
<input type="checkbox"/> Oily	<input type="checkbox"/> Other:

Last Rainfall	
<input type="checkbox"/> More than a week ago	<input type="checkbox"/> During the last week
<input checked="" type="checkbox"/> During the last 24 hours	<input type="checkbox"/> Raining Now
Amount of rain:	10 mm

Drains		
Drain Flow:	<input type="checkbox"/> Drain flowing	<input checked="" type="checkbox"/> Drain not flowing